



Certificate of Compliance

Certificate: 80033019

Master Contract: 213862

Project: 80033019

Date Issued: 2020-02-18

Issued To: Vaisala Oyj
PO Box 26
Helsinki, ., 00421
Finland

Attention: J-P Mörsky

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Sam Zaffino*
Sam Zaffino



PRODUCTS

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT Certified to US Standards

Model: MI Multichannel User Interface; rated 100-240Vac, 50/60Hz, 60VA (AC version) or 24Vdc, 60W (DC version); two sets of relay contacts per module card rated 3.0A max at 150Vac, all other inputs and outputs SELV, Ambient 0 to 50°C, Pollution degree 4, Installation Category II, Enclosure Type 4X.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

ANSI/UL Std. No. 61010-1 (3rd)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 80033019

Master Contract: 213862

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80033019	2020-02-18	Transfer from 70061842 and to cover alternate construction